

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 01 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130578</u> ✓
Name <u>Malik Jones</u>		ID# <u>5830</u> ✓ Date <u>6/26/10</u>
A	Agency <u>AST</u>	Phone # <u>(907) 451-5100</u>
Instrument Location <u>Cantwell</u>		
Alco S/N <u>58442</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>1225</u>		
B	Alco Test Value Average <u>.070</u> ✓ 1 <sup>st</sup> Alco	
	<u>.071</u> ✓ 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>		
(OVER)		

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15<sup>th</sup> day of APRIL, 2011

[Signature]  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130578

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY

DATAMASTER cdm 130578

JUNE 26, 2010

OPERATOR'S NAME:

JONES/MALIK/JAHAL

OPERATOR'S NUMBER: 5830

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

V

O.L. #: 0

DEPT/AGENCY: CNTD

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: 58442

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 27.00 in

ALCO TARGET .074 13:51

BLANK TEST .000 13:52

INTERNAL STANDARD VERIFIED 13:52

ALCO TV 27.00 in .070 13:53

BLANK TEST .000 13:54

SUBJECT SAMPLE .000 13:54

BLANK TEST .000 13:55

ALCO TV 27.00 in .071 13:56

BLANK TEST .000 13:56

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY

DATAMASTER cdm 130578

JUNE 26, 2010

TIME 14:01

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 51c

BREATH TUBE: 41c

BAROMETER: 27.80 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~